

Customer: John R Adam & Sons
 Address: King George V Dock
 Renfrew Road
 Govan
 Glasgow
 G51 4SD

Certificate Number:	700163
Location:	Warehouse
Equipment Type:	Platform
Manufacturer:	Avery Weigh-Tronix
Model:	E1070
Serial Number:	081150169
Capacity:	3000.0 kg
Division:	1.0 kg
Permissible Error:	+/- 3.0 kg

Test Equipment	
Equipment Type:	Weights
Set Number/Serial Number:	FC01-FC52
Certificate Number:	W18060
Certificate Expiry Date:	13/03/2019

Repeatability Test			
Applied Load:		2000.0 kg	
As Found Test		Definitive Test NA	
	Error		Error
1	2000.0 kg 0g	1	
2	2000.0 kg 0g	2	
3	2000.0 kg 0g	3	

Comments:
 No issues.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
0g	0g	0g	0g
500.0 kg	500.0 kg	500.0 kg	0g
1000.0 kg	1000.0 kg	1000.0 kg	0g
1500.0 kg	1500.0 kg	1500.0 kg	0g
2000.0 kg	2000.0 kg	2000.0 kg	0g
3000.0 kg	3000.0 kg	3000.0 kg	0g
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test							
Applied Load:		2000.0 kg					
	As Found Test	Error	Definitive Test NA				
		Error					
1	2000.0 kg	0g	1				
2	2000.0 kg	0g	2				
3	2000.0 kg	0g	3				
4	2000.0 kg	0g	4				
Scale Layout:		<table border="1"> <tr> <td>4</td> <td>1</td> </tr> <tr> <td>3</td> <td>2</td> </tr> </table>		4	1	3	2
4	1						
3	2						

IND

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis
 Signature:

Date: 06 August 2018

Customers Name: Walter Bennett
 Signature:

Date: 06 August 2018