



Customer: John R Adam & Sons  
Address: King George V Dock  
Renfrew Road  
Govan  
Glasgow  
G51 4SD

Certificate Number: 700162  
Location: Warehouse  
Equipment Type: Platform  
Manufacturer: Avery Weigh-Tronix  
Model: ZM201  
Serial Number: 162550096  
Capacity: 3000.0 kg  
Division: 1.0 kg  
Permissible Error: +/- 3.0 kg

**Test Equipment**  
Equipment Type: Weights  
Set Number/Serial Number: FC01-FC52  
Certificate Number: W18060  
Certificate Expiry Date: 13/03/2019

Repeatability Test			
Applied Load:		2000.0 kg	
As Found Test		Definitive Test NA	
	Error		Error
1	2000.0 kg	0g	1
2	2000.0 kg	0g	2
3	2000.0 kg	0g	3

Comments:  
No issues.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
0g	0g	0g	0g
500.0 kg	500.0 kg	500.0 kg	0g
1000.0 kg	1000.0 kg	1000.0 kg	0g
1500.0 kg	1500.0 kg	1500.0 kg	0g
2000.0 kg	2000.0 kg	2000.0 kg	0g
3000.0 kg	3000.0 kg	3000.0 kg	0g
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test			
Applied Load:		2000.0 kg	
As Found Test	Error	Definitive Test NA	Error
1	2000.0 kg	0g	1
2	2000.0 kg	0g	2
3	2000.0 kg	0g	3
4	2000.0 kg	0g	4
Scale Layout:		4	1
		3	2

IND

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis  
Signature: *Barry Ellis*  
Date: 06 August 2018

Customers Name: Walter Bennett  
Signature: *Walter Bennett*  
Date: 06 August 2018

