



Customer: John R Adam & Sons
Address: King George V Dock
Renfrew Road
Govan
Glasgow
G51 4SD

Certificate Number: 700162
Location: Warehouse
Equipment Type: Platform
Manufacturer: Avery Weigh-Tronix
Model: ZM201
Serial Number: 162550096
Capacity: 3000.0 kg
Division: 1.0 kg
Permissible Error: +/- 3.0 kg

Test Equipment
Equipment Type: Weights
Set Number/Serial Number: FC01-FC52
Certificate Number: W18060
Certificate Expiry Date: 13/03/2019

Repeatability Test			
Applied Load:		2000.0 kg	
As Found Test		Definitive Test NA	
	Error		Error
1	2000.0 kg	0g	1
2	2000.0 kg	0g	2
3	2000.0 kg	0g	3

Comments:
No issues.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
0g	0g	0g	0g
500.0 kg	500.0 kg	500.0 kg	0g
1000.0 kg	1000.0 kg	1000.0 kg	0g
1500.0 kg	1500.0 kg	1500.0 kg	0g
2000.0 kg	2000.0 kg	2000.0 kg	0g
3000.0 kg	3000.0 kg	3000.0 kg	0g
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test			
Applied Load:		2000.0 kg	
As Found Test	Error	Definitive Test NA	Error
1	2000.0 kg	0g	1
2	2000.0 kg	0g	2
3	2000.0 kg	0g	3
4	2000.0 kg	0g	4
Scale Layout:	4	1	
	3	2	

IND

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis
Signature: *Barry Ellis*
Date: 06 August 2018

Customers Name: Walter Bennett
Signature: *Walter Bennett*
Date: 06 August 2018

