

Calibration Report Number: 700161

Company: John R Adam & Sons  
 Address: King George V Dock  
 Renfrew Road  
 Glasgow  
 G51 4SD

Weighbridge Type: Surface  
 Weighbridge Size: 18m x 3m  
 DWI Manufacturer: Shering  
 DWI Model:: Command 940  
 DWI Serial No: 3153646  
 Capacity (kg): 50,000  
 Division (kg): 20  
 Test Weights Used: FC01-FC52  
 Cert Number: W18060  
 Calibration Date: 14/03/2018  
 Permissible Error:  
 0kg-20,000kg=+/-20kg, 20,000kg-40,000kg=+/-40kg,  
 40,000kg-50,000kg=+/-60kg  
 Next Cal Date:

As Found Check


Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
20,000	20,000	20,000	0
24,000	24,000	24,000	0
Substitute Load	18,720	18,720	0
28,720	28,720	28,720	0
33,720	33,720	33,720	0
38,720	38,720	38,720	0
Substitute Load + FLT + 24000	47,500	47,500	0

Forklift + Driver Weight (kg): 4780

Repeatability Check Test Weight (kg) = 47,500

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	47,500	0
2	0	0	47,500	0
3	0	0	47,500	0

Weighbridge Corner Layout



Comments: No issues

Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,780	6,780	0
2	6,780	6,780	0
3	6,780	6,780	0
4	6,780	6,780	0
5	6,780	6,780	0
6	6,780	6,780	0

Engineer: Barry Ellis  
 Signature:

Customer: Walter Bennett  
 Signature:

Date: 06 August 2018 Date: 06 August 2018